

March, 29th 2021

Declaration of conformity

Herewith the ACE Handels- und Entwicklungs GmbH (ACE Instruments) seated in 83395 Freilassing / Germany assures that all ACE mouthpieces fulfill a high hygienic standard.

This means that each ACE mouthpiece is hygienically single packed and has a non-return valve. Such a valve prevents accidental aspiration of saliva or germs from the injection channel of alcohol testers.

Please note: The ACE mouthpieces are for single use only. This means that per person and per test always a new mouthpiece should be used. The single packaging guarantees that you will always have a hygienically perfect mouthpiece.

ACE Instruments informs that you must not disinfect, wash out or clean and reuse an already used mouthpiece. ACE Instruments disclaims any liability for damages resulting from the repeated use of one and the same mouthpiece.



According to the test report number PB2009598 issued on December, 11th 2020 by DEKRA Automobil GmbH, Labor für Umwelt- und Produktanalytik, Handwerkstr. 17, 70565 Stuttgart, Germany, no odour and taste deviation of the simulant food was detected during sensory testing (level 0).

According to Article 3 paragraph 1 c of Regulation 1935/2004, materials and articles shall be manufactured in such a way that, under normal or foreseeable conditions of use, they do not transfer constituents to food in quantities liable to affect the organoleptic characteristics of the food. The ACE mouthpieces fulfil these conditions.

The limit value for global migration according to Article 12 (1) of Regulation (EU) 10/2011 of 10 mg/dm² was complied with for the 1st migrate for 10 % ethanol.

In addition, the tested sample was saliva- and sweat-proof.


ACE Handels- & Entwicklungs GmbH
Staufenstraße 1 · 83395 Freilassing / Germany
info@ace-technik.com · www.ace-technik.com
T +49 (0) 8654 / 588 38 0 · F +49 (0) 8654 / 588 38 98

Company stamp / Signature